


Quantitative Imaging for Colocalization Analysis

Spectroscopy,
not Photography

Daniel J. White



CBG
Max Planck Institute
of Molecular Cell Biology
and Genetics



Topics:

- ✓ Images = “Information”

(Digital Images)

- ✓ Limitations of microscopy hardware

- ✓ Colocalization analysis

Coloc_2

What is an Image anyway..?


- 1st: Go to Basics course slides

Spatial Digitisation


Experimental Design - First Think...

Quantitative Experiments?


- Am I trying to measure the size/shape of some type of object(s)



- Am I trying to see movement over time?





- Am I trying to measure a number, amount or concentration?



Am I trying to measure the number of some type of object?

- Can I define how my objects appear in images?
- Segmentation
 - Image intensity - threshold
 - Size - threshold
 - Shape - circularity etc.



Am I trying to see something move over time?

- Can I define what movement is?

- Linear - A to B?

- Direction

- Speed


- Velocity



- Rotation




- Clustering




Am I trying to measure an amount or concentration?

- Does that have a Biological meaning?
- Absolute or Relative?
- Can I calibrate my image intensity vs. something else / itself?
 - eg. Fluorescence signal vs. Quantitative Assay or Baseline, Control
 - Fluorescence response might not be linear!



Am I trying to measure an “image parameter”?

- Does that have a Biological meaning?
- Absolute or Relative?
 - Total / Mean / SD of signal
 - Background
 - Signal : Noise
 - Texture (smooth/spotty)
 - “Colocalization” between “colours” / channels




● Hardware issues

- Optical problems
- Spectral problems
- Intensity Digitisation
- Colour Channels

http://fiji.sc/Colocalization_-_hardware_setup_and_image_acquisition


Avoid Emission Bleed Through and Crosstalk/Cross-excitation

- Dye selection / Filter selection
 - Emission bleed through and/or excitation crosstalk...
 - Means you get: Overlapping emission - Quantitative? No!



Use multi tracking (Zeiss) / sequential (Olympus)



Beware ! Crosstalk and Bleed Through



Cross talk (wrong excitation) **Bleed through (wrong emission)**

Watch Out - More Holes To Fall Into:



- Correct objective lens / microscope setup for task
- N.A / Resolution.
- Apochromat for different colours (UV)
- Calibrate Scanner / Check with multi-colour beads



Check with multi-colour beads



● Widefield: (Dvcore 1 micron tetraspek beads):

- Optimise Filter alignment / angle
- Lenses have residual aberrations, even expensive ones.



Check with multi-colour beads

- Confocal (Zeiss 510):
 - Calibrate Scanner + Align pinholes (and collimator)
- Measure error – then, correct for it!




Watch Out - More Holes To Fall Into:


- Required bit depth - 8 bit often enough for LSCM imaging... and colocalization analysis.
- More bits only for quantitative experiments where small intensity differences are measured.
- 12 bit - bigger files than 8 bit.
(Olympus... 12 bit only. Zeiss 8,12. Leica 8,12,16.)
- 16 bit file is 2x bigger in RAM / on disk, than 8 bit !
- CCD - many cases 12 bit might give better coloc info.

Watch Out - More Holes To Fall Into:

- Laser power - don't bleach area before imaging it.
 - Bleached sample
 - Lower signal : noise
 - Lost information
 - Set the HV and Offset quickly (Auto HV)
 - Live imaging, bleaching - big problem
Use low laser power (but more noise)



Colocalization / Correlation



The past:

“I see yellow - therefore there is colocalization”

but published images “look” over exposed.


No colocalization definition + No stats = No Science.

From Now On: 3D. Quantification. Correlation. Statistics.

Complementary methods: BioChemical, Optical (FRET, FLIM)

Colour Merge Images? Only for Art!

- Channel Merge Images? What are they good for?
 - Apart from looking pretty... not much.
 - Scientific conclusions from the image below?
 - Colour blind people - see green and red the same!
- Use Magenta / Green or Yellow / Blue




Colour Merge + Projection = Danger!


Never make colour merge / overlay images from projections of 3D / z stacks... why not?

Lose 3D info - are the objects overlapping in 3D, or is one in front of the other one, in the z-stack.

False overlaps!!! Easy to make false interpretation



colour merged projection



3 D

What does “Colocalisation” mean anyway...?

- 🌐 That depends who you ask...
- 🌐 ... and what **BIOLOGY** you are thinking about



+



=



Colocalisation/Correlation?


Think about the biology!

- What is the biological/biochemical question?
 - Are you looking for Co-Compartmentalisation?
 - Are you looking for exclusion / anti correlation?
 - Are you looking for interacting molecules?
 - Then you must also do biochemistry
(Immuno Co-precip, Fluo Correlation Spectroscopy)
 - FRET / FLIM might be very informative


Colocalisation / Correlation / Concurrence?

“Colocalisation” covers two qualitatively different conditions:

1) that objects have both
fluorophores present
(Object Based Coloc)
Segmentation needed.
Biology?



2) there is some relationship
between the **intensities** of
the fluorophores in a pixel.
(Pixel Intensity Based Coloc)
Interaction - BioChemistry?



Colocalization / Correlation / Concurrence?

2 fluorophores are there in a pixel


Binary information

Is it Random?

Is it Real?

Little or no biological meaning?

...unless you are confident about
how to segment objects out from
the background.




Definition of Terms

- “Concurrence” = “co-presence” “there is red and green”
- “Colocalisation” = Relationship between channel intensities
 - Eg. “Red is only found with Green”
- Special case - “Correlation”
 - Intensity Correlation over Space

Define what is Colocalization / Correlation?

Colocalisation is #1




2 objects overlap
Binary information
No intensity information

Concurrence?
Image Segmentation!

Biological Meaning?

Colocalisation is #2




Some objects appear to
overlap
with another object
Binary information
No intensity information

Colocalisation?

Biological Meaning?

Colocalisation is: #3



X

Intensity profiles overlap




Image “Correlation”


Biological Meaning?
Co-compartmentalisation?
Physical interaction?

Colocalisation/Correlation -Think about:


- Are your “objects” smaller than optical resolution?
 - Vesicles? Small Organelles?
 - Check channel overlap with sub resolution beads!
- Are your objects large?
 - Large single homogenous blobs?
 - Large reticular networks / membranes
 - Resolution required?
- Complementary “correlation” methods
 - Fluorescence correlation spectroscopy (FCS in live cells)
 - Flow Cytometry? Multiple markers in a cell. Good stats.

Colour Merge Images = Bad ... so what should I do instead?


- “Colocalisation Analysis”
- Statistical Significance of Colocalisation
 - Single image - random / insignificant.
 - Statistical P value (significance), Manders coefficients, and Scatter Plot. (ImageJ, BioImageXD, Huygens and others)
- But remember...
 - Don't merge projections of stacks (you lose 3D info, false coloc)
 - Don't believe your eyes, they lie. Machines don't make mistakes...



Colocalization Analysis




vs.




How can I measure the amount of colocalisation or rather “correlation” between these two images?

BioImageXD, ImageJ and others have methods to do that!



Colocalization Analysis



Scatter plot
2D histogram
Publish it?

Coloc stats:
Pearsons r
 M_1 , M_2 ,
Costes P-val,

Automatic
thresholding

Pearson's Image Correlation Coefficient (Manders et al., 1993)

$$r = \frac{\sum_i (R_i - R_{av}) \cdot (G_i - G_{av})}{\sqrt{\sum_i (R_i - R_{av})^2 \cdot (G_i - G_{av})^2}}$$

Don't panic - it's not that complicated!

Correlation between images, r ranges from -1 to +1

+1 means full correlation (images are the same)

0 means no correlation (random)

-1 means full anti correlation (no red where there is green)

Pearson's Image Correlation Coefficient

In English...per pixel and summed for the whole image:

$$r = \frac{\text{sum of (red intensity - average red) x (green intensity - average green)}}{\text{sqrt of squares of above}}$$


$r = +1$




$r = -1$



$r = 0$




$r > 0$



Pearson's Image Correlation Coefficient is...

- Insensitive to diff. intensity of the 2 images. Why?
- Insensitive to intensity offset.
- If red is 1/2 as bright as green...
 - Still can get $r = 1$
 - ... so Pearson's r is robust for biological imaging...



Manders' Coefficients

$$M_R = \frac{\sum_i R_{i,\text{coloc}}}{\sum_i R_{i,\text{total}}}$$

$$M_G = \frac{\sum_i G_{i,\text{coloc}}}{\sum_i G_{i,\text{total}}}$$

Biologically meaningful
coloc coefficients:


Proportion of each dye
colocalised with the other
(Manders et al., 1993)

$R_{i,\text{coloc}}$ = colocalized red signal

$R_{i,\text{total}}$ = total red signal

Great! ... but how do I know which pixels are
colocalized and which are not...?

“Thresholding” and “% colocalisation”



The calculated
“% colocalisation”
depends on what
thresholds you set.


... so how should
one set them?

..until you get the
result you want?

No science here!



Automatic Thresholding?

- How should I set the thresholds of the 2 channels?
 - Manually? No! Subjective user bias, not reproducible...
 - Need a robust reproducible method!
 - Find thresholds where Pearson correlation below thresholds ≤ 0



2D Histograms / Scatterplots

- Display 2 colour channel image data in 2D:
 - colour merge / overlay or 2D histogram?
 - 2D histogram: Ch1 - y axis (left), Ch2 - x axis (bottom)
 - Colour mapped to number of pixels with that R and G value (right)




Green


Red

Overlay


Complete

**A**


Diff. intensities

**B**

Partial

**C**

Exclusion

**D**






Fig. 5. Colocalization analysis with JACoP; Pearson and Manders, scatter plots and correlation coefficients. Scatter plots (A–D) correspond to the

2D Histograms / Scatterplots


- See correlation qualitatively - better than colour merge
- See problems from imaging:




Saturated
Noisy







Saturated
No correlation?



Wrong offset




Wrong offset
Bleed through



Automatic Thresholding?


- Does it work in a biological experiment? Yes!
- Time course of Rev-CRM1 dissociation, nucleolus to nucleus
- The dissociation rate constant $k_d = 1.25 \pm 0.31 \times 10^{-3} \text{ s}^{-1}$




One more thing...

● Statistical significance!

- Are coloc results better than random chance?
- A busy image might give high correlation and Manders
- Lots of signal = larger chance of random signal overlap.



vs.




17 / 40 pixels
overlap !!!

Is that significant
or just random?


Costes' Method - Randomisation...

- Measure Pearson's correlation for:
 - Randomised 1st channel image data (PSF sized chunks)
 - Repeat 100 times
 - How many randomised have \leq correlation than real image.
 - If $> 95\%$ of randomised are worse, then we believe Manders.

$P = 0.5 = 50\%$ (no)
 $P = 0.95 = 95\%$ (yes)
 $P = 1 = 100\%$ (YES!)
confidence





vs.



Colocalization example: **virus** entry to **caveolae**



10 min P.I.



32% of virus colocalized

Costes P-value 0.00
0% chance it's real

20 min P.I.



39% of virus colocalized

Costes P-value 1.00
100% chance it's real

Without significance test, we wrongly assume virus is colocalised
with caveolae at 10 min P.I.

It is not! Only at 20 min is there significant correlation.



Examples: No Correlation?

Pearson r 0.024

M1 0.0354

M2 0.0471

Why high
Thresholds?



Noisy Saturated Images

Good Correlation?





Pearson r 0.747

M1 0.7291

M2 0.7420

Thresholds
Include
noise?

Badly
Saturated!



Bad detector settings Good Correlation?





Pearson r 0.68

M1 0.77


M2 0.63

Offset wrong
+ Saturated

Thresholds
Handle it?
No?




Bleed Through! DAPI into GFP




Bad detector settings

Good Correlation? Bleed through?



Bad detector settings... ...gives wrong results!!!



Software for Colocalization

ImageJ - Colocalization plugins


- Coloc_2, JACoP, older plugins.

BioImageXD (Coloc Task - Pixel Intensity and Object based methods)

Huygens (RBNCC)

Imaris (Coloc module)

Matlab (J-Y. Tinevez, MPI-CBG / Pasteur)



Thanks to: MPI-CBG LMF and IPF
Fiji, Heino, Pahajoki,
Kankaanpää, Marjomäki
Uuksalainen, Paavolainen,
TEKES, Tom Kazimiers

Thanks for listening